

Test certificate No.: MDV - 11/2019 - 005

CERTIFICATE OF ANALYSIS OF FINISHED PRODUCT
STANDARD : ISO 4074 - 2015 (E)

Batch No : QU19002
Exp date : 09/2024

Quantity : 172.800 PCS

Brand : QUINN 0.01 LUBRICATED ULTRATHIN (PACK 12) Customer : Medevice Korea

	Dimensions			Individual packs with visibly open seals	Visible defects	Freedom from Holes	Bursting volume (Unaged)	Bursting pressure (Unaged)	Package seal Integrity	Lubricant Qty	Individual packaging material marking
	Length	Width	Thickness								
Units	mm	mm	mm	*	*	*	dm3	KPa	*	mg / Pc	*
Specified Value	Min 180 mm	53± 2	0.058-0.065	No open seals	*	*	Min 18.0	Min 1.0	No leaks	450 ÷ 550	*
Tested Value	Range	188.0 - 191.0	0.058 - 0.060	N.A	N.A	N.A	12.4 - 48.7	0.75 - 1.85	N.A	470 - 510	N.A
	Mean	189.8	0.059	N.A	N.A	N.A	39.4	1.54	N.A	500	N.A
	Defects / Rejects (Nos.)	0	0	0	0	0	1	3	0	0	0
Sample Size	13 Pcs	13 Pcs	13 Pcs	315 pcs	315 Pcs	315 Pcs	315 Pcs	315 Pcs	32 Pcs	13 Pcs	13 Pcs
AQL	NA	NA	NA	0.4	0.4	0.25	1.5	1.5	2.5	NA	NA
Ac	0	0	0	3	3	2	10	10	2	0	0
Re	1	1	1	4	4	3	11	11	3	1	1
Accepted / Rejected	Accepted	Accepted	Accepted	Accepted	Accepted	Accepted	Accepted	Accepted	Accepted	Accepted	Accepted

Total Sample Size : 688 Pcs

Testing date : 04/11/2019

QA Inspection Checked By :

QA Inspectors Signature

CÔNG TY LIÊN DOANH MEDEVICE 3S

QC

Ngày .4. tháng .11. năm 2019

Nguyễn Thị Hồng

Phạm Thị Ngọc Quỳnh